















Testing and Validation Methodology



- Verify the assumption that there exists a certain correlation between the Peak-to-Peak periods of the different Cardiovascular signals.
- Challenges
 - Transmission delay
 - Hidalgo ECG (Chest), Vernier ECG (Arm), and Sp_{o2} (Finger)
 Sampling Frequencies
 - Hidalgo ECG (256 Hz), Vernier ECG (200 Hz), Sp_{O2} (75 Hz)
 Up/ down sampling
- Mathodology
 Collect simultaneous data offline from all three devices
 MATLAB function

 Period estimation
 Hidalgo data: Downs ampling to 200 Hz
 Sp₀₂ data: Up sampling to 200 Hz

Testing and Validation

Results of Testing and Validation



- Standard Deviation / Standard Deviation of Difference
 Minimum and Maximum Value
- Threshold
- Percentage Error

%	Error =	Avg(Mean of	Difference)	²) × 100%	
		Signal	Mean	× 100%	

	Sn	Manualan		
	3P02	vernier	Hidalgo	
Sp ₀₂	0	0.0057	0.0057	0.765%
Vernier		0	0.0023	0.537%
Hidalgo			0	0.537%

